



Annual Compliance and Performance Report

2021

Best Thertronics Ltd.
413 March Road
Ottawa, Ontario, Canada
K2K 0E4

Class 1B License

NSPFL-14.00/2029

Reporting Period: January 1st, 2021 to December 31st, 2021

| Report Submitted, Rev 0: March 30, 2022
Rev 1:

Report Revision History

Revision	Submission Date	Comments
0	March 30, 2022	<ul style="list-style-type: none">• Original submission

Table of Contents

1	Introduction	5
1.1	Compliance with Other Regulatory Agencies	6
1.2	New Licensed Activities.....	6
1.3	Significant Modifications or Changes to Site or Facility.....	6
	Safety and Control Areas	6
1.4	Management System.....	6
1.5	Human Performance Management	9
1.6	Operating Performance	10
1.7	Safety Analysis	12
1.8	Physical Design.....	12
1.9	Fitness for Service	13
1.10	Radiation Protection	13
1.11	Conventional Health & Safety.....	16
1.12	Environmental Protection.....	18
1.13	Emergency Management and Fire Protection	18
1.14	Waste Management	20
1.15	Security	23
1.16	Safeguards and Non-proliferation	23
1.17	Packaging and Transport.....	23
2	Other Matters of Regulatory Interest.....	24
2.1	Licensee's Public Information and Disclosure Program	24
2.2	Financial Guarantees.....	24
3	Concluding Remarks.....	25
3.1	Signing Authority Certification.....	25
	Appendix A – Lost Time Statistics	26

This page has been left blank intentionally.

1 Introduction

Best Theratronics Ltd. (BTL) is a medical device manufacturing company, located at 413 March road, of medical equipment used throughout the world. The main products that require the possession of a Class 1B licence include:

- Cobalt 60-based external beam radiation therapy units,
- Cesium 137-based self-contained irradiators (SCIs) for blood or research irradiation,
- Cyclotrons with beam energies ranging from 6 to 70 MeV.

In September of 2018, a renewal application was submitted to the CNSC for a period of 10 years, until June 30, 2029. Best Theratronics was granted a renewed Class 1B license on July 1, 2019.

Licence NSPFL-14.00/2029 authorizes Best Theratronics to:

- a) operate a Class 1B nuclear facility located at 413 March Road, Ottawa, Ontario, including activities related to:
 - i. operating a particle accelerator/accelerators (cyclotron/cyclotrons);
 - ii. possessing nuclear substances for the purposes of manufacturing radiation devices and radioactive source teletherapy machines;
 - iii. possessing a radioactive source teletherapy machine, for the purposes of developing and testing
- b) possess, transfer, manage, and store nuclear substances arising from the activities regarding the particle accelerators;
- c) produce prescribed equipment;
- d) possess, transfer, use, import, export, manage, and store within the facility any nuclear substances that are required for, associated with, or related to manufacturing radiation devices, and development and testing of radioactive source teletherapy machines;
- e) possess, transfer, use, import, export, and store prescribed equipment that is required for, associated with, or related to manufacturing of radiation devices and development and testing of radioactive source teletherapy machines, and manufacturing radioactive source teletherapy machines; and
- f) possess and use prescribed information that is required for, associated with, or arise from operating the Class 1B nuclear facility.

In addition to the Class 1B nuclear substance and processing facility licence, Best Theratronics possesses two other CNSC licences (Class II Nuclear Facilities and Prescribed Equipment Licence & Nuclear Substances and Radiation Devices Licence) in order to conduct service work on prescribed equipment sold to customers within Canada. Information related to these activities is reported in their respective Annual Compliance Reports (ACRs).

This ACR is submitted with respect to licence condition 3.2 and reflect information related to the NSPFL-14.00/2029 activities.

1.1 Compliance with Other Regulatory Agencies

In manufacturing medical devices that are sold and shipped internationally, Best Theratronics is required to comply with many standards and regulatory agencies. Compliance is required by agencies such as:

- International Organization for Standardization (ISO 13485, ISO 9001)
- Health Canada
- United States Nuclear Regulatory Commission (US NRC)
- Federal Drug Administration (FDA)
- United States Department of Transportation (US DOT)
- Medical Directive of Europe
- Other international regulatory agencies where Best Theratronics devices are sold

Within Canada, Best Theratronics complies with all federal, provincial, and municipal regulations in order to operate. Oversight agencies include:

- Transport Canada – Transportation of Dangerous Goods (TDG) Regulations
- Canada Occupational Health and Safety Regulations
- Ministry of Environment (National Pollutant Release Inventory)
- Ministry of the Environment and Climate Change (Hazardous Waste Information Network)

1.2 New Licensed Activities

No operational changes occurred in 2021. There were no new Class 1B licensed activities since the last compliance monitoring period.

1.3 Significant Modifications or Changes to Site or Facility

A small package delivery door with a security camera was installed in shipping and receiving in April 20, 2021. Its purpose is to limit access to BTL's shipping bay. Only large semi-trucks with pre-authorization are allowed to load/unload in the shipping bay.

Safety and Control Areas

1.4 Management System

1.4.1 Applicable Activities

Best Theratronics is committed to developing, manufacturing, installing and servicing safe and quality products and to continually improve the effectiveness of the quality management system to meet customer and regulatory requirements for health care and research products and services.

The quality management system is applicable to all Best Theratronics CNSC licensed activities. Best Theratronics has established several management systems to help guarantee this commitment. These management systems include:

- Training, Personnel Examination and Certification

- Work Organization
- Fitness for Duty of Personnel and Facilities
- Procedure Documentation
- Culture of Safety and Compliance

The implementations of these management systems are discussed in the following safety and control areas (SCA) sections in this report. As a manufacturing facility of medical devices, the overall management system implemented follows current ISO standards.

1.4.2 Management System Effectiveness

Compliance to Best Theratronics' CNSC licence conditions are assessed in-house in the areas of security, emergency management and fire response, waste management, environmental protection, and radiation protection. Refer to the following SCA sections for more information.

Management review team (MRT) meetings are conducted annually to analyze and discuss general trends of the organization. Best Theratronics held one large Management Review Team meeting in 2021 for the operations over 2020. Throughout the year many small, informal meetings were held. The following topics were discussed:

- Quality policy
- Environmental, Health & Safety Policy
- Quality, Environment and Health & Safety Objectives
- Audits
- Post Market Surveillance
- Process Performance and Product Conformity
- Status of Corrective and Preventative Actions and OFI's
- Follow-up Actions from Previous Management Reviews
- Changes that could affect the quality management system or the organization structure
- Effectiveness of Actions Taken to address Risks and Opportunities
- Recommendations for Improvement
- New or Revised Regulatory Requirements
- Review of Risk Methodology
- Self-Assessments of Management Processes
- Safety culture
- Radiation Control Program
- Trend Analysis
- Best Theratronics Training Plan

The overall quality system and objectives were discussed, reviewing the quality system to ensure that each objective remains applicable and effective. Some action items were created to improve the evaluation of the quality system, which will be followed in the next MRT meeting.

1.4.2.1 Annual Quality Management System Audit

Best Theratronics completes an annual internal audit of the overall quality management system. The scope of the audit covers the review of company objectives, policies and procedures, the management standard, requirements of ISO13485:2016, ISO 9001:2015, and the Medical Device Single Audit Program (MDSAP). An annual internal audit was completed in April, 2021.

In 2020, Best Theratronics started transitioning to an alternate ISO auditing/notified body. A certification audit was completed by GMED in June, 2021.

1.4.2.2 CNSC Management Systems Inspection

The CNSC conducted a virtual desktop inspection on Management Systems in November 2020, resulting in 2 notices of non-compliance. Best Theratronics opened CAPAs which were addressed and closed in 2021.

1.4.2.3 Organizational Structure for the Management and Control of Licensed Activities

Minor changes to Best Theratronics' organizational structure occurred in 2021. The position of Director of Manufacturing & Facilities Operations became vacant. These director duties were split between the Production Planning Manager the supply chain department. The following roles outline the personnel employed to ensure licensed activities are properly managed at Best Theratronics:

- Quality & Regulatory Manager
- Director of Engineering
- Director of Cyclotron Operations
- Technical Services Manager
- Production Planning Manager
- Supply Chain Manager
- Radiation Safety Officer
- Radiation Safety Specialist
- Radiometric Measurement Specialist
- Medical Physicist
- Compliance Specialist
- Quality Control Supervisor
- Production Supervisors (3)
- Contract Security Supervisor
- Contract Security Officers

1.4.3 Document Changes

Below is a list of the documents that were updated in 2021. Updates to such documents reflect changes in regulation, audit observations, and corrective action implementation. Updated versions of documents supporting the Class 1B licence were submitted to the CNSC as per requirement in the Licence Conditions Handbook:

- 3.24-AA-01 *Design Change Procedure*

Revised to indicate the three different scenarios of full validation for a design change.

- C1B-SD-15b *Public Information*
Reviewed and revised Section 3.6, Table 1 – target disclosure timeframes.
- 5.08-RP-09 Leak Test Sampling and Measurement
Revised leak testing frequency and procedure for sealed sources.

1.5 Human Performance Management

Best Theratronics has implemented a robust human performance management system that ensures that staff is sufficient in numbers and have the required knowledge, skills and training to safely carry out their duties. Staff levels are monitored by supervisors and managers to ensure there is sufficient personnel. Regular meetings between the Directors and the President are also used to assess staffing levels.

Elements of a Systematic Approach to Training (SAT) have been implemented for positions where *the consequence of human error poses a risk to the environment, the health and safety of persons, or to the security of the nuclear facilities and of nuclear substances*”. If ever an employee’s roles or responsibilities change, their training requirements are reviewed.

1.5.1 Training Programs

At Best Theratronics various environmental health and safety training programs have been implemented to ensure safe working environments for all employees. Upon employment employees are trained on Best Theratronics’ policies regarding compliance, security, environmental impacts and the quality system expectations. The following table lists the environmental health and safety training programs are conducted at Best Theratronics.

On an annual basis, radiation safety refresher training is required for Class II service technicians to ensure safe practices are applied at customer’s sites within Canada and internationally.

Table 1: Training programs offered at BTL and frequency that refresher training is mandated.

Training Program	Refresher Frequency
Chemical Spill	3 years
Crane	3 years
Emergency Response	2 years
First-Aid	3 years
Fork-lift/Pallet Truck	3 years
Lead Control	3 years
WHMIS	3 years
Nuclear Energy Worker/Radiation Safety	3 years
Nuclear Energy Worker Service Refresher	1 years
Transportation of Dangerous Goods	2 years

1.5.2 Training Effectiveness Evaluation

The training program at Best Theratronics is evaluated through:

- On-the-job training assessment by the trainer
- Review of CAPAs that indicate a root cause linked to inadequate training
- Regular trend analysis on key indicator quality systems processes

- Training evaluation forms following in-class instructor training

For training courses that have a graded learning assessment in order for completion, a grade of at least 70% must be achieved to pass the course. The following table identifies the number of employees trained in 2021.

Table 2: Number of personnel trained in 2021 for each training program offered at BTL.

Training	# of personnel trained in 2021
Crane	27
First-Aid	3
Fork-lift/Pallet Truck	14
Lead control	10
WHMIS	28
Transportation of Dangerous Goods	20
Nuclear Energy Worker/Radiation Safety Nuclear Energy Worker Service Refresher	57

All personnel trained in 2021 successfully passed the end of course evaluations. Review of SAT-based training programs is complete, which re-analyzed training requirements, training techniques, and assessed the incorporation of additional training modules.

1.5.2.1 Radiation Safety Training

During the reporting year, fifty seven employees successfully completed Nuclear Energy Worker radiation safety training. This grouping includes facility personnel who required initial NEW training and refresher training, in addition to Best Theratronics' service personnel, who complete radiation safety refresher training annually. In 2021, no radiation related incidences occurred where the root cause was determined to be due to lack of training.

1.5.2.2 CNSC Human Performance Management Inspection

The CNSC conducted a remote inspection on Human Performance Management in November 2020, resulting in 4 notices of non-compliance and 3 recommendations in response to the implementation of the Systematic Approach to Training Method. Best Theratronics opened CAPAs and implemented changes to the affected areas and policies in 2021, with the final outstanding response being completed in January, 2022.

1.5.3 Sufficient Number of Qualified Workers

Management Review Team discussions are conducted to ensure that there are an appropriate amount of qualified personnel to continue operations in a safe manner. Best Theratronics has security personnel on-site at all times. An emergency contact list is available and tested twice per year, ensuring upper management and appropriate response personnel are reachable.

1.6 Operating Performance

As an ISO 9001:2015 certified facility, Best Theratronics operating performance program integrates operating experience, adequacy of procedures, and the conduct of licensed activities.

Operating Experience is evaluated using a Corrective Action Preventative Action (CAPA) system, capturing non-conformances and improvement opportunities discovered through audits. Reporting and trending of operational experiences are discussed at the annual MRT meeting and monthly Health & Safety meetings. Concerns regarding licensed activities are discussed within Radiation Safety & Security Committee meetings occurring quarterly. Weekly meetings regarding radioactive material shipments are conducted with members of the radioactive production team, regulatory compliance, logistics, and customer service. Email notification updates are sent out to key operational members of the organization to keep all those involved informed and to track notifications sent to the CNSC.

Procedures are reviewed, updated and implemented on a regular basis to align with revised regulations. Training on updated procedures takes the form of *Self-Study Review* where all training is coordinated and maintained by their training coordinators.

1.6.1 Licensed Activities Audits Overview

The CNSC conducted one audit in 2021 on the following topics:

- Packaging and transport

The audit found that BTL had a robust program that included appropriate training, certification, transport documentation, and labeling.

1.6.2 Reportable Events

On Nov. 5, 2021 BTL shipped four loaded GammaCells to SwRI, Texas, USA for long-term storage or disposal without a Canadian export license. BTL self-reported the incident on Feb. 5 2022. Details can be found in CAPA220301, which was issued and closed, revising policy 5.03-AA-07 ‘Licenses to export risk-significant radioactive sources’, and retraining customer service, logistics, and regulatory employees.

1.6.1 Operational Limits

The basis of obtaining the Class 1B License for the Best Theratronics facility was to manufacture and test Class II prescribed equipment and cyclotrons for the medical and research industries. In 2021, Best Theratronics operated within the limits outlined in the Class 1B license.

1.6.1.1 Class II Workload

The R&D Class II prescribed equipment located in Cell 4 (T1000, S/N 4) was operated for a total 47 hours, where all hours were related to research. Operational information is provided in the following table.

Table 3: Operational information for Class II prescribed equipment located in Cell 4.

Source Serial Number	Source Type	Beam On Time [hrs]	Output at 1m [Gy/min]	Output date reference	Output at 1m used for analysis [Gy/min]	Total work load (Gy)
S-6245	Co-60	32	1.37	January 1, 2021	1.33	2548
S-6306	Co-60	15	1.17	January 1, 2021	1.03	930

<i>Beam on total [hrs]</i>	47	<i>Total work load [Gy]</i>	3478
----------------------------	----	-----------------------------	------

1.6.1.2 Cyclotron Operations

The operating limits stated in Best Theratronics Licence Conditions Handbook are related to cyclotron development and testing. Operational information is provided in the following table.

Table 4: Operational information for Class 1B prescribed equipment located in High Bay.

Model/Serial number	B6P01	B15P02	B35P01	B70P02
Nominal Beam energy	6MeV	15MeV	35MeV	70MeV
Operating beam energy	<1MeV in factory for FAT	<1MeV in factory	<1MeV in factory	<1MeV in factory
Beam current (nominal)	500μA nominal 550μA tested	500μA nominal	1200μA nominal	1000μA nominal
Beam-on times	FAT test 2 hours Total test time ≈ 8 hours	The cyclotron is in manufacturing stage	The cyclotron is in manufacturing stage	The cyclotron is in manufacturing stage

1.7 Safety Analysis

Safety analysis reports are undertaken prior to design and implementation of changes to critical safety components, including devices, transport containers, and facilities. Safety analysis reports are reviewed by the management review team.

Overall workplace safety is monitored by two committees in order to maintain the safe and healthy occupational working environments. The Workplace Health & Safety Committee is responsible for monitoring operations and recommends improvements to management. Radiation-related safety concerns are discussed in meetings held by the Radiation Safety & Security Committee.

1.7.1 Facility Safety Improvements

The facility is toured and inspected by two members of the Health & Safety Committee on a monthly basis. There were no significant Health & Safety concerns brought up from these inspections.

1.8 Physical Design

A design change process for the control, management, evaluation, release, completion and implementation of changes to Best Theratronics drawings and documents is implemented. In 2020, Best Theratronics applied for a CNSC Class II Facility construction licence to prepare a pre-existing bunker for development and testing of Class II equipment. Modification plans were submitted to the CNSC for review in April 2020, and a pre-construction inspection was completed in November 2020. The development and testing work associated with this project progresses under the design change process.

1.9 Fitness for Service

1.9.1 Effectiveness of Maintenance and Testing Programs

Best Theratronics maintains an inventory of radiation survey meters, radiation area monitors, and personal digital reading dosimeters. Monthly checks of these instruments are completed to ensure all radiation monitoring equipment are in good working condition and not past their calibration due dates. In 2021, all required equipment were maintained and made available in good working order. In the event that operational deficiencies were discovered, immediate repairs were completed to prevent potential health and safety issues, or units were retired if repairs were deemed to not be cost effective.

Preventative maintenance on production equipment is performed at regularly scheduled intervals determined by the usage, operation history, and manufacturers' recommendations where available. Maintenance schedules are maintained for each piece of equipment and are reviewed quarterly for completeness. In 2021, there were no issues related to the operation of any of the manufacturing equipment.

In addition, Best Theratronics assesses its facility on an on-going basis through monthly Health & Safety inspections, general review of the facility and as concerns are presented from employees.

1.9.2 Effectiveness of Aging Management Strategies

Best Theratronics Facilities & Maintenance team assesses the requirement for upgrades to existing machinery and improvements required around the facility. A representative from the Facilities & Maintenance team is a member of the Health & Safety Committee and is actively involved in aging management discussions, providing first hand information to management.

To improve the versatility of the manufacturing shop machining capabilities at the facility, a new sandblasting booth was purchased in 2021 and installed to replace an aging unit. It has not been put into service as of this date but a 2022 launch is planned.

1.10 Radiation Protection

1.10.1 ALARA Principle Application

Adherence to the application of the *As Low As Reasonably Achievable* (ALARA) principle within Best Theratronics is supported by the main tenants of training, monitoring employee radiation exposure, and planning for special work. Initial Nuclear Energy Worker (NEW) training is provided and a refresher course is mandatory every 3 years to maintain the NEW status and radiological awareness. NEWs are designated based on their work tasks, required controlled area access, and the likelihood of receiving a higher dose than the public annual effective dose limit of 1 mSv. Personal doses of NEWs are monitored, on either monthly or quarterly basis, with the use of personal dosimeters alongside recorded doses from electronic personal dosimeters (EPDs). In addition, area monitors are installed throughout the facility to alarm if radiation fields exceed normal levels. A special work permit system, requiring authorization by the RSO, is implemented. This system identifies any special work that falls outside of normal, routine work to ensure it is properly planned to minimize unnecessary radiation exposures. Radiation protection

assessments, consisting of monitoring for contamination and radiation surveys, are completed monthly to ensure ALARA doses in both controlled and accessible areas.

The Radiation Safety & Security Committee (RSSC) meets regularly to review radiation-related safety matters at Best Theratronics. The meetings take place to discuss concerns and identify improvements to the overall safety and security culture at Best Theratronics. In 2021, quarterly meetings were held to ensure effective communication of radiation-related work and security concerns.

1.10.1.1 ALARA Action Level Reportable Incidences

There were no Class 1B ALARA Action Level exceedences in 2021.

1.10.2 Radiation Protection Program Performance

Following an audit on the Radiation Protection Program in 2016, administrative levels of effective and equivalent doses were decreased to provide a better indication of the application of the program. In addition, in-house wipe test and surface contamination trigger levels were reduced. These levels were decreased to better reflect current operations. In 2021, there were no incidents where radiation exposure action levels were exceeded.

No other radiation related events occurred in 2021. The radiation protection training program has proven to be robust.

1.10.3 Radiation Protection Program Improvements

On an annual basis, an internal audit of the radiation protection program is conducted. The internal audit for 2021 revealed that minor administrative updates were required in order to reflect current operations and procedure clarification. Continued improvements to the radiation protection program include:

- In-house training improvements by encouraging communication with all staff to help identify gaps in training related to licensed activity tasks
- Providing additional dose monitoring badges for new or infrequent radiation related work
- Incorporate additional practical trainings for SAT-based tasks
- Discussing industry operational experience within the Radiation Safety and Security Meetings

1.10.4 Dose Monitoring Data

All individuals requiring access to controlled areas where radioactive material is stored, in addition to completing work where they may exceed the public annual dose limit of 1 mSv, are classified as a NEWS. Only NEWS are allowed in such areas and are monitored with the use of personal dosimeters as part of the Radiation Protection Program. Doses are monitored for two groups of NEWS at Best Theratronics:

- 1) Device Manufacturing and Class II Research and Development Employees (Building Personnel)
- 2) Class II Servicing Employees

Group 1 employees are reported under the Class 1B License. Class II Servicing Employee doses are reported with the respective Class II Servicing Licenses (14127-3-28/14127-8-24). On occasion, qualified Class II Servicing employees participate in Class 1B licensed activities. All NEW doses associated with Class 1B work is reported in this section as Class 1B NEWS.

Extremity monitoring is applied to NEWs whose job tasks require working with their hands in close proximity to radioactive material, such as service technicians or radiation device welders. Workers are required to wear two extremity Optically Stimulated Luminescent Dosimeters (OSLD), one on each hand. The distribution of occupationally obtained doses is listed in the following table for both effective and extremity doses.

Table 5: Effective and extremity radiation dose distribution for Class 1B NEW employees at BTL.

Work Group	Total Monitored	Dose Range (mSv)					
		<0.01	0.01-1.00	1.01-5.00	5.01-10.00	10.01-20.00	>20.01
Effective Dose		<0.01	0.01-1.00	1.01-5.00	5.01-10.00	10.01-20.00	>20.01
Class 1B NEWs	64	53	11	0	0	0	0
Class II Servicing	<i>Reported in Class II servicing licenses ACRs (14127-3-28/14127-8-24)</i>						
Extremity (Maximum between hands)		<0.01	0.01-1.00	1.01-5.00	5.01-10.00	10.01-20.00	>20.01
Class 1B NEWs	16	14	2	0	0	0	0
Class II Servicing	<i>Reported in Class II servicing licenses ACRs (14127-3-28/14127-8-24)</i>						

Best Theratronics operates with occupational doses below the maximum allowable NEW effective dose of 50 mSv in one dosimetry year and 500 mSv per year for extremities. The following table provides the dose data for 2021:

Table 6: Dose statistics for Class 1B NEW employees at BTL.

2021 Class 1B NEWS	Effective	Extremity
Total workers monitored	64	16
Collective dose (mSv)	0.75	0.93
Average dose , with zeros (mSv)	0.01	0.06
Average dose, measured only (mSv)	0.02	0.47
Maximum dose received (mSv)	0.13	0.47

The following table provide Class 1B NEW dose data from 2016 – 2021.

Table 7: Average and maximum values of effective and extremity doses for Class 1B NEW employees at BTL between 2016-2021.

Class 1B NEW Effective Doses							
	2016	2017	2018	2019	2020	2021	Regulatory Limit
Total workers monitored	60	68	68	68	73	64	--
Average dose , with zeros (mSv)	0.03	0.02	0.16	0.04	0.01	0.01	--
Maximum dose received (mSv)	0.98	0.47	8.65	1.00	0.19	0.13	50 mSv
Class 1B NEW Extremity Doses							
	2016	2017	2018	2019	2020	2021	Regulatory Limit
Total workers monitored	17	16	18	19	19	16	--
Average dose , with zeros (mSv)	0.09	0.07	1.41	0.22	0.15	0.06	--
Maximum dose received (mSv)	1.10	0.50	13.51	2.51	2.4	0.47	500 mSv

The number of NEWS monitored over the past six years has remained consistent. The average effective dose over this period has fluctuated between 0.01-0.04 mSv, with the exception of 2018. The maximum effective doses received have fluctuated between 0.13 - 1.00 mSv. For extremity doses, the average extremity dose has fluctuated between 0.06-0.15 mSv, with the exception of the maximum value in 2018. This single incident resulted in an effective dose and an extremity dose action level exceedence for two personnel conducting Class IB licensed tasks. Maximum doses received conducting Class IB licensed activities at Best Theratronics over the past six years account for 17% and 3% of the regulatory limits, for effective dose and extremity dose respectively.

1.10.5 Routine Radiation Protection Assessments

Best Theratronics conducts monthly checks in areas of the facility likely to show signs of radiological contamination or increased radiation fields for both controlled and uncontrolled areas. Internal monitoring limits for radiation fields are 1 mR/h for controlled areas and 0.1 mR/h in uncontrolled areas. All monthly facility surveys were found to be within these limits throughout the monitoring period. No abnormal readings were found in 2021.

Areas within the facility where radioactive material is stored or transported are checked for signs of contamination on a monthly basis. Contamination checks are also performed on an as-needed basis; from incoming radioactive shipments to the movement of depleted uranium inventory around the facility. All facility contamination checks were within acceptable limits and no incidences were found where radioactive contamination was of concern. No contamination events occurred in 2021.

When radioactive shipments are received at Best Theratronics, the radiation field is measured to ensure the packages are within the Transport of Dangerous Goods Regulations. Additionally, all receipts that are intended to contain radioactive sources are wiped for surface contamination to ensure contamination events are isolated prior to unloading. No incidences where transport package radiation surveys exceeded regulatory limits were observed or package surface contamination were detected in 2021. A CNSC remote radiation protection audit was scheduled and completed in January of 2022.

1.11 Conventional Health & Safety

The Best Theratronics Health and Safety Program is centered around prevention, first aid, investigations, hazardous substance awareness, an employee's right to refuse dangerous work acknowledgement, and workplace inspections.

1.11.1 Conventional Health & Safety Committee

The Health & Safety Committee members are responsible for reviewing reports on the investigations of occupational injuries, hazardous occurrences and near misses. The Best Theratronics Health and Safety Committee met on 10 occasions during 2021. Health and safety audits of the facility were also conducted monthly; all findings were made action items, and recorded in the meeting minutes. Twenty six (26) action items opened in 2021, as well as one item from 2020, were completed and closed in 2021.

1.11.1.1 Conventional Health & Safety Program Improvements

As a result of workplace observations and concerns discussed within the Health & Safety Committee meetings, the following areas of improvement were tracked in 2021:

- Chemical spill response program
- Lead control program
- Increased salting of walkways and parking lot
- Ensuring emergency exits are cleared
- Lead pouring ergonomics
- Safety needs communication
- Tripping hazards and the maintenance of painted safety lines on the manufacturing shop floor
- Review of required PPE for sandblasting
- Transportation of gas cylinders within the facility
- Operation of a new phone and paging system
- Development of a common machine shop H&S checklist
- Development of COVID-19 safety protocols

1.11.1.2 Health & Safety Occurrences

In 2021, Best Theratronics documented a total of 9 medical reports, 1 of which required outside medical attention, and 1 near miss report. These incidents caused cuts, bruises, nausea/vomiting, numbness, and a slip on ice. The following graph shows a breakdown of the health and safety reports, including lost time incidences.




Figure 1: Distribution of significance of incurred workplace injuries between 2016-2021.

In all instances, medical reports were reviewed and corrective actions were introduced if appropriate. Workplace injuries and lost time incidences are reviewed on a monthly basis by the Health and Safety Committee to ensure effort is put forth to prevent future occurrences.

A Lost Time Incident (LTI) occurs if an employee suffers a workplace injury resulting in an absence from work past the day of the incident, loss of wages, or a permanent disability/impairment. There were zero

Lost Time Incidences in 2021. There was one potential LTI in 2021 where medical treatment was sought and the employee returned to work.

- April 25, 2021 – Worker experienced whole body numbness and was transported via ambulance to the hospital. No long term effects occurred and the worker returned to work the next day as normal. It was determined that the worker's condition was unrelated to the workplace environment.

Lost time frequency and severity rate are provided in Appendix A for reference between 2016-2021.

1.11.1.3 COVID-19 Health and Safety Committee

In addition to the 10 Health & Safety Committee meetings per calendar year, the Health & Safety Committee members met on a monthly basis to discuss issues related to employee safety as it pertains to COVID-19.

1.12 Environmental Protection

An emissions analysis was completed in 2013 in support of an Environmental Compliance Approval (Air) application. This analysis assumed all significant emission sources were operating simultaneously at their individual maximum rates of production. The results indicated that manufacturing operation emission concentrations are below regulatory limits, demonstrating Best Theratronics' compliance with O. Reg. 419/05: Air Pollution – Local Air Quality.

Best Theratronics manufacturing operations do not produce airborne or liquid radiological releases to the environment as on-site sources are double encapsulated by a third party. The radioactive material used in Best Theratronics manufactured medical devices is contained within a welded stainless steel encapsulation. The loaded transport container or loaded self-shielded irradiators are stored within a radiation designated area within the facility. All radioactive materials are double encapsulated sealed sources or depleted uranium, therefore there are no releases into the environment and do not pose as an exposure hazard to the public.

All aspects of Best Theratronics' operations that may have an impact on the environment are identified, evaluated, recorded and reviewed periodically.

Operations of the facility do not produce airborne or liquid radiological emissions. No environmental releases occurred in 2021.

1.13 Emergency Management and Fire Protection

As a manufacturing facility for medical devices, where radioactive sources are stored on site, fire and radiological emergency programs are required to ensure the safety of Best Theratronics.

1.13.1 Emergency Preparedness

Aspects of Best Theratronics' Emergency Response Program are tested periodically, as indicated in the following table.

Table 8: Emergency Preparedness Test/Drills

Emergency Test/Drill	Minimum Testing Frequency
Emergency Personnel Call List	Semi-Annually
Fire Evacuation Alarm and Drill	Annually
Fire Alarms	Annually
Radiation Alarms	Monthly, Quarterly (Klaxon)
Emergency Power	Monthly
Full scale evacuation exercise*	Once every five years
First aid casualty (<i>as part of First Aid training</i>)	Every three years
Chemical Spill	Periodically
Communication test for equipment and effectiveness	Periodically (everyday use)

*Full scale evacuation last occurred in April 2019.

1.13.2 Program Effectiveness

The Emergency Response Committee (ERC) meets at least once a year to oversee emergency response planning at Best Theratronics Ltd. The last meeting took place on September 24, 2021. The following action items were discussed:

- 1) Emergency response procedural updates – review methods of incorporating an attendance system for building evacuations
- 2) Improvements to emergency response training – three training opportunities were found to increase safety awareness
- 3) Emergency preparedness resource inventory – introduce reference resources to communicate emergency procedures for all personnel on the premises
- 4) Communication and identification resources – introduce Emergency Response Committee identification vests and communication devices

A review of Emergency Response procedures was conducted, as well as worker refresher training regarding ER, to ensure Best Theratronics is adequately prepared to respond in an emergency situation.

1.13.3 Fire Protection Program Performance

Best Theratronics has implemented various measures to improve fire safety at the workplace. Elements of the fire protection program at Best Theratronics include:

- a hot work program
- developed combustibles policy
- refresher training of flammables and combustible liquids
- fire warden training
- training on the correct use of electrical cords

Routine checks of all fire protection related equipment are conducted, at a frequency listed in the following table, to ensure functionality when required.

Table 9: Life safety equipment testing frequency.

Equipment	Testing Frequency
Fire Alarm System	Monthly
Emergency Lighting	Monthly
Fire Extinguishers	Monthly
Sprinklers	Quarterly

1.13.4 Fire Protection Program Effectiveness

The fire protection program effectiveness was assessed during the fire drill on November 9, 2021. The main improvement point was to remind workers to use the closest emergency exit during an evacuation.

1.14 Waste Management

To reduce the impact on the environment, Best Theratronics has established a waste management program to promote the safe handling and disposal of waste generated from its operations.

1.14.1 Non-Radioactive Hazardous Materials

The landfill waste stream of 27.3 MT in 2021 dropped compared to the values from previous years. This decreased value is attributed to a general drop in waste removed from the building and an increase in recyclables.

Table 10: Waste (in MT) that is disposed of into each waste stream between 2016-2021.

Waste Stream	2016 (MT)	2017 (MT)	2018 (MT)	2019 (MT)	2020 (MT)	2021 (MT)
Waste to Landfill	34	21.6	39.5	38.22	34.75	27.3
Recycled Paper, Cardboard and Shredded Paper	20	20	18*	25.6*	24	25.7
Recycled Glass, Aluminum Cans & Plastics	0.7	0.7	0.3*	1.5*	1.3	1.0
Recycled Metal	39	7.7	12.13	33.8	24.8	27.3
Other Recovered Material	5.6	5.6	5.6	5.63	5.63	5.98
Totals	99.3	55.6	75.53*	104.82*	90.48	87.3
Diversion Rate	66%	61%	48%*	63.5%*	62%	69%

*Upon review of 2018 and 2019 data for the 2020 Waste Audit Report, these values have been corrected.

The following recommendations are keyed to the largest components of the landfill waste stream:

- Continue to search out options to divert wood waste from landfill.

- Implement a program to collect food waste and paper towels from and send them to a composting facility rather than landfill. Collectively, food waste and paper towels account for about 61% of landfill waste.
- Review the materials entering the 20 yd³ construction waste bin and assess if any of these materials can be diverted from landfill.
- Continue to support and strengthen the use of existing recycling programs through communications and review of bin placement to optimize employee participation.

Best Theratronics' hazardous waste management program is responsible for the proper disposal wastes such as chemical waste, electronics, paint, batteries, construction/demolition waste, and PCB containing light ballasts and fluorescent light bulbs. The following table provides the amounts of hazardous waste removed between 2016 and 2020.

Table 11: Hazardous waste (in kg) that was disposed of between 2016-2020.

Waste Code	Description	2016	2017	2018	2019	2020	2021
112	Lead acid batteries	5 kg	--	--	--	--	--
114C		--	--	--	--	--	10 L
122	Alkaline batteries	--	--	70 kg	--	--	--
145	Paint	--	--	--	5 kg	80 L	--
146	Filters with lead dust	48 kg	--	--	--	--	--
146	Florescent bulbs and HID lamps	150 kg	140 kg	65 kg	185 kg	20 kg	--
146	Lead contaminated material	--	--	--	70 kg	460 kg	--
146	Zirconium alloy scrap	225 kg	--	--	800 kg	800 kg	--
148	Inorganic acid oxidizer	88 L	--	16 L	--	--	--
148	Corrosive liquid (nitric acid, sodium chloride)	--	--	--	--	48 L	45 L
212	Acetone	1015 L	600 L	530 L	620 L	620 L	--
212	Antifreeze	--	--	--	10 L	--	--
212	Glycol/water	--	--	--	--	140 L	--
251	Watery oil	--	--	--	12900 L	900 L	2000 L
252	PCB ballasts	--	40 kg	10 kg	--	--	--
252	Machine oil	1965 L	1980 L	1920 L	--	1220 L	--
263	Organic flammable waste	200 L	245 L	340 L	56 L	60 L	200 L
331	Organic gas aerosols	8 L	20 kg	73 Kg	16 kg	60 kg	40 kg
253	Emulsified oil	--	--	1000 L	--	--	--

1.14.2 Radioactive Hazardous Materials

In order to be compliant with ISO 14001:2015, Best Theratronics revised its environmental management system to include the identification and evaluation of operations that may have an impact on the environment on an annual basis. A number of environmental objectives have previously been determined and tracked by the MRT throughout the licensing period. They include:

- Dispose of, or transfer, sealed sources at 413 March road to a licensed facility.
- Dispose of, or transfer, prescribed equipment containing radioactive source to a licensed facility.

- Dispose of, or transfer, depleted uranium at 413 March Road to a licensed facility.

Best Theratronics has an end-of-life management program for the exhausted and returned sealed sources. These sources are sorted into three categories upon return: for reuse, for transfer, or for disposal.

- Sources that are destined for reuse, include sources for re-encapsulation, to be incorporated into Best Theratronics self-contained irradiators or teletherapy machines.
- Sources that are transferred to other manufacturers for recycling. These sources are shipped to other suppliers or manufacturers of Co-60 sources, where the capsules will be cut open and the radioactive material reused in the manufacturing of new sources for other purposes.
- Sources that are destined for disposal are transferred to licensed disposal facilities, such as Canadian Nuclear Laboratories, for long-term storage and eventual disposal.

In 2021, a total source activity of 1347 TBq was diverted from the disposal stream and reused or recycled according to Best Theratronics' end-of-life management program. The graph below indicates the activity breakdown of the managed sources between 2016-2021. Values above each year indicate the total activity managed in the respective year.




Figure 2: Radioactive waste (TBq) in each of the reuse, transfer, or disposal waste management streams.

The depleted uranium inventory at Best Theratronics originates from returned components of legacy teletherapy units and other legacy items. This inventory is temporarily stored at Best Theratronics awaiting proper disposal through the end of life management program. No disposal or recycling of depleted uranium occurred in 2021. Best Theratronics is actively seeking opportunities to recycle the depleted uranium.

1.15 Security

1.15.1 Site Security

Best Theratronics has an adequate security program in place, where the site-security plan is reviewed on a regular basis. Concerns regarding the security of radioactive material are discussed on a regular basis, during Radiation Safety & Security Committee (RSSC) meetings.

No security-related events occurred in 2021.

1.15.2 Transport Security

Limited and approved carriers of radioactive material are contracted to further ensure the security of devices or components containing radioactive material during transit. These carriers are audited annually to ensure their procedures comply with current regulations and Best Theratronics' security policies. Transportation security plans of the radioactive material carriers were found to be adequate.

No transport security reportable incidences occurred in 2021.

1.15.3 Personnel Security

As part of Best Theratronics' employment process, all employees are required to supply a criminal's records check at the start of their employment. Best Theratronics has implemented a criminal record check renewal policy every five years. This policy has been fully implemented.

1.16 Safeguards and Non-proliferation

1.16.1 Safeguards and Non-proliferation Program Performance

Best Theratronics possess and temporarily stores depleted uranium from legacy teletherapy units destined for disposal. Accounting and reporting of Best Theratronics' inventory of depleted uranium and other materials containing depleted uranium are completed annually as per REGDOC-2.13.1 *Safeguards and Nuclear Material Accountancy*.

The annual Physical Inventory Taking (PIT) produced no discrepancies between the physical values and the reported values to the CNSC. Best Theratronics was not selected for either the CNSC Physical Inventory Taking – Evaluation (PIT-E) or the IAEA Physical Inventory Verification (PIV) in 2021.

1.17 Packaging and Transport

Best Theratronics prepares, packages and ships medical devices containing sealed Category 1 and 2 radioactive materials worldwide. The Packing and Transport program at Best Theratronics meets the requirements of the CNSC *Packaging and Transport of Nuclear Substances Regulations* (2015), IAEA SSR-6 (2012), Transport Canada *Transportation of Dangerous Goods*, USDOT 49 CFR, and US NRC 10 CFR.

Radioactive sealed source shipments are transported in Type A or certified Type B containers. Best Theratronics implements a transport container maintenance and inspection program in accordance with IAEA SSR-6 2012. In addition to annual inspections, containers undergo a routine inspection each time they are returned from the field.

2 Other Matters of Regulatory Interest

2.1 Licensee's Public Information and Disclosure Program

2.1.1 Public Inquiries and Media Coverage

The public is encouraged to contact Best Theratronics for more information regarding concerns through the info@theraronics.ca email address available on the Best Theratronics website. There were no public inquiries received in 2021. Best Theratronics understands the importance of Indigenous relations, and as such, has previously reached out to surrounding Indigenous groups including the Algonquins of Ontario (AOO). In 2021, Best Theratronics received no direct inquiries from Indigenous groups.

As per Best Theratronics' obligation to keep the public informed, the Best Theratronics website is updated with information for public inquiry. The updates to the website include:

- Annual compliance reports (ACRs) for all of Best Theratronics' CNSC licences (servicing and Class 1B)
- Notifications of licence renewals
- Annual reports on lead (and its compounds)
- Notification of false alarms and building evacuations
- Incidents occurred where any reporting or action level was exceeded

There were a total of four (4) updates to the public information program pages in 2021.

2.1.2 Facility Tours

Due to the COVID-19 pandemic, Best Theratronics held no tours of the facility in 2021.

2.1.3 Future Public Information Program Plans

Best Theratronics reviewed its Public Information and Disclosure program in 2021. Minor program updates were made to improve feedback and communication with the target audience. Best Theratronics will continue to monitor its public information program performance. Best Theratronics plans to continue hosting public information sessions and facility tours once safe to do so.

2.2 Financial Guarantees

As of July 2017, Best Theratronics has estimated decommissioning costs to be \$1.80 million. This includes a 25% contingency amount. In 2021, Best Theratronics removed 2041 TBq of source activity from its possession. These sources were either reused (935 TBq), transferred to another licensed facility to be recycled (411 TBq), or disposed of (694 TBq).

Best Theratronics currently has in place the total amount of the financial guarantee with the CNSC in the amount of \$1.8 million. This is in support of Best Theratronics' current licenses. This financial guarantee is in the form of a Letter of Credit, issued by Canadian Banks.

The financial guarantee will be maintained on a continuing basis. As the decommissioning plan is revised, due to on-going decommissioning activities or changes to the operational program, the Letter of

Credit will also be revised to ensure sufficiency to fund decommissioning activities. The next full review of the financial guarantee will take place in 2022.

3 Concluding Remarks

The Class 1B license offers Best Theratronics increased flexibility in its operations. Despite this, Best Theratronics operating status in 2021 did not change significantly from previous years. There were no major events, observations, or non-compliances identified during 2021 that would affect the safety and security of personnel, the public, or the environment.

Best Theratronics continues to make adequate provisions for the protection of the environment and the safety of both employees and the public. Best Theratronics acts in compliance with the licensing conditions set out in license NSPFL-14.00-2029 and the associated Licensing Conditions Handbook.

3.1 Signing Authority Certification

I hereby certify that Best Theratronics has been operating in compliance with license NSPFL-14.00/2029, except where otherwise noted.



Matthew Efseaff, PhD, MSc

Radiation Safety Officer

613-591-2100 ext 2762

Appendix A – Lost Time Statistics

Table 13: Lost Time and Frequency Rate calculated for 2016-2021.

	2016	2017	2018	2019	2020	2021
# of LTIs	3	1	2	2	0	0
Frequency Rate	2.05	0.684	1.37	1.37	0.00	0.00
Total Missed Days	55	22	12	8	0	0
Severity Rate	37.6	15.0	8.21	5.470	0.00	0.00

*Assumption is 150 employees working 37.5 hrs/week for 52 weeks.